

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005 & ANSI/NCSL Z540-1-1994

MATPHIL TECHNOLOGIES, INC. DBA ACCUTEK LABORATORIES DBA PIPETTE.COM

10360 Sorrento Valley Rd, Suite E San Diego, CA 92121 Alex Spector Phone: 858 536 3344

CALIBRATION

Valid To: November 30, 2017 Certificate Number: 2902.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Fluid Quantities

Parameter/Equipment	Range	CMC ^{2, 3} (±)	Comments
Volume	0.2 μL 1 μL 2 μL 5 μL 10 μL 20 μL 50 μL 100 μL 200 μL 500 μL 1000 μL 2000 μL 2000 μL 2000 μL 2000 μL 2000 μL 2000 μL 2000 μL	0.01 μL 0.01 μL 0.02 μL 0.05 μL 0.05 μL 0.21 μL 0.30 μL 0.70 μL 0.85 μL 3.0 μL 3.0 μL 2.3 μL 3.9 μL 6.7 μL 13 μL	5-, 6- and 7-place ultra micro balances.

¹ This laboratory offers commercial calibration service.

A- (. But

A- (. But

² Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of k = 2.8. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ The technique involves gravimetric pipetting of laboratory grade water (reference material) obtained from approved preferred vendor, Fisher Scientific.



Accredited Laboratory

A2LA has accredited

MATPHIL TECHNOLOGIES, INC. DBA ACCUTEK LABORATORIES DBA PIPETTE.COM

San Diego, CA

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 General requirements for the competence of testing and calibration laboratories. This laboratory also meets the requirements of ANSI/NCSLI Z540-1-1994 and any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009).

FOR LABORATOR TOCREOUS AND THE SEAL SEAL SEAL SEAL TO SEAL TO

Presented this 8th day of January 2016.

Senior Director, Quality & Communications

For the Accreditation Council

Certificate Number 2902.01

Valid to November 30, 2017

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.